

The University of New South Wales

Optics & Radiometry Laboratory

School of Optometry and Vision Science

COMMERCIAL-IN-CONFIDENCE

LAMP CIRCUIT POWER MEASUREMENTS

Report Number 16041.2

for

**Nikkon Lighting Pty. Ltd.
Unit 3, 30-32 Artisan Road
Seven Hills NSW 2147**

Testing completed: 23 February 2016
Report issued: 23 February 2016

*Optics & Radiometry Laboratory
School of Optometry and Vision Science
Room LG22, Old Main Building (K15)
Gate 14, Barker Street
UNSW SYDNEY NSW 2052
Phone: +61 2 9385 4622
Fax: +61 2 9313 8602
Email: orlab@unsw.edu.au*

This document shall not be reproduced, except in full



Accreditation No. 1923

The results of tests, calibrations and/or measurements included in this document are traceable to Australian/national standards.

NATA is a signatory to the ILAC mutual recognition arrangement for the mutual recognition of the equivalence of testing, calibration and inspection reports

Accredited for compliance with ISO/IEC 17025.

Page 1 of 3

Checked by


A handwritten signature in black ink, appearing to be 'mc', is written next to the 'Checked by' text.

LAMP CIRCUIT POWER MEASUREMENTS

Description of sample

ORLAB Sample ID	Make	Model	Control Gear	Description
16041-2	Nikon Lighting	S2125-125W	PHILIPS 9137108590	LED floodlight, 125 watts nominal with clear lens energised with integral Philips (150W 0.70A) LED driver.
16041-3				
16041-4		S2110-125W		

NB. The customer has advised that the 2 models listed are electrically identical. Only model S2125-125W was tested.



This document shall not be reproduced, except in full



Accreditation No. 1923

Accredited for compliance with ISO/IEC 17025.

The results of tests, calibrations and/or measurements included in this document are traceable to Australian/national standards.

NATA is a signatory to the ILAC mutual recognition arrangement for the mutual recognition of the equivalence of testing, calibration and inspection reports

Page 2 of 3

Checked by *me*

Equipment used

ID	Description	Calibration due
237	Hewlett Packard 34401A Digital multimeter as a volt and frequency meter	15-Feb-17
123	Fluke 45 Digital multimeter as an ammeter	02-Feb-17
311	Hameg Powermeter HM8115-2	19-Oct-16

Temperature

Ambient temperature 25.1 °C

Uncertainties of measurement

The uncertainty values stated in this report have been calculated at the 95% level of confidence ($k=2$). This means that the chance that a result is in error by more than the value of its uncertainty is not more than five in one hundred.

Results

Voltage 240.0 \pm 0.5 VAC
Frequency 50.00 \pm 0.02 Hz

Sample ID	Current AAC	Power WAC	Power factor
16041-2	0.559 \pm 0.002	131.9 \pm 0.5	0.98 \pm 0.01
16041-3	0.560 \pm 0.002	132.1 \pm 0.5	0.98 \pm 0.01
16041-4	0.559 \pm 0.002	132.5 \pm 0.5	0.99 \pm 0.01
AVERAGE	0.559 \pm 0.002	132.2 \pm 0.5	0.98 \pm 0.01

Voltage 250.0 \pm 0.5 VAC
Frequency 50.00 \pm 0.02 Hz

Sample ID	Current AAC	Power WAC	Power factor
16041-2	0.538 \pm 0.002	131.9 \pm 0.5	0.98 \pm 0.01
16041-3	0.539 \pm 0.002	132.2 \pm 0.5	0.98 \pm 0.01
16041-4	0.538 \pm 0.002	132.1 \pm 0.5	0.98 \pm 0.01
AVERAGE	0.538 \pm 0.002	132.1 \pm 0.5	0.98 \pm 0.01

Stephen V Dain

Stephen Dain
Authorised Signatory

This document shall not be reproduced, except in full

Accredited for compliance with ISO/IEC 17025.

Page 3 of 3

Checked by *jl*



Accreditation No. 1923

The results of tests, calibrations and/or measurements included in this document are traceable to Australian/national standards.

NATA is a signatory to the ILAC mutual recognition arrangement for the mutual recognition of the equivalence of testing, calibration and inspection reports